

## Certificate of Specification

---

**Material :** 351695T                      PAPANICOLAOU'S EA 50 STANDARDSTAIN  
**Batch :**

---

**CHARACTERISTICS****SPECIFICATIONS**

---

Absorption maximum I 1	630 - 634 nm
Absorption maximum I 2	530 - 534 nm
Absorbance A1, I1 (I <sup>m</sup> <sup>ax</sup> ; diluted 1:200	0.120 - 0.300
Absorbance A2, I2 (I <sup>m</sup> <sup>ax</sup> ; diluted 1:200	1.900 - 2.300
Suitability for microscopy (Vaginal smea	passes test

We certify that this batch conforms to the specifications listed above.  
BDL : Below detected limit.

Dr. Wolfgang KnicheI responsible laboratory manager quality control  
VWR International  
Document printed on 27.03.2013

*This document has been produced electronically and is valid without a signature.*